

IAF/IAA SPACE LIFE SCIENCES SYMPOSIUM (A1)  
Human Physiology in Space (2)

Author: Dr. Arturo Moleti  
Università di Roma "Tor Vergata", Italy

Dr. Triestino Minniti  
University of Rome "Tor Vergata", Italy

Dr. Altea Russo  
University of Rome "Tor Vergata", Italy

Dr. Yoshita Sharma  
University of Rome - Tor Vergata, Italy

Dr. Andrea Civiero  
Università degli studi di Trieste, Italy

Dr. Renata Sisto  
INAIL Research, Italy

Dr. Filippo Sanjust  
Italy

Dr. Teresa Botti  
INAIL Research, Italy

Dr. Luigi Cerini  
Italy

Mr. Giorgio Pennazza  
Università Campus Bio-Medico di Roma, Italy

Dr. Marco Santonico  
Italy

Dr. Alessandro Zompanti  
Italy

Mrs. Rosa Sapone  
Altec S.p.A., Italy

Mr. Alessandro Crisafi  
ALTEC Spa, Italy

Mr. Maurizio Deffacis  
Altec S.p.A., Italy

Dr. Marco Lucertini  
Italian Air Force , Italy

Dr. Dario Castagnolo  
Telespazio S.p.A., Italy

Dr. Gabriele Mascetti  
Italian Space Agency (ASI), Italy

Mr. Giovanni Valentini  
Italian Space Agency (ASI), Italy

Dr. Monia Vadrucchi  
Italian Space Agency (ASI), Italy

Prof.Dr. Maria Patrizia Orlando

ADVANCED AUDIOLOGICAL TOOLS FOR NON-INVASIVE MONITOR OF INTRACRANIAL  
PRESSURE IN MICROGRAVITY

**Abstract**

Otoacoustic emissions (OAEs) are known to provide a non-invasive indirect indicator of intracranial pressure (ICP) changes. Indeed, the OAE phase is sensitive to the middle ear transmission, which depends on ICP due to its connection with the intracochlear pressure. Increased middle ear stiffness is associated with increased OAE phase. Advanced OAE acquisition procedures based on the equalization of the forward pressure level (FPL) of the stimuli add independent information about the middle ear transmission function that may complement and validate the OAE phase measurements. FPL techniques require preliminary evaluation of the Thevenin parameters (pressure and impedance) of the sound sources. When the Thevenin calibrated otoacoustic probe is inserted in the ear canal, a direct estimate of the load impedance of the ear canal (essentially dependent on the middle ear transmission) is obtained. A decrease of the imaginary part of the load impedance in the low-frequency range is the unambiguous signature of the increased middle ear stiffness that is assumed to generate also the OAE phase change when ICP increases. Thus, the simultaneous occurrence of the two phenomena (increased OAE phase and decreased imaginary part of the impedance) validates the diagnostic interpretation of OAE changes as due to ICP changes. Acoustic Diagnostics is an Italian Space Agency experiment operating on the International Space Station 2019-2022. Five ISS astronauts were audiotologically tested, measuring distortion product OAEs (DPOAEs) pre-, in-, and post-flight, to detect changes in their hearing properties associated with long term exposure to microgravity conditions. The payload measured DPOAE level and phase using FPL calibration, and Thevenin calibrated sound sources. Therefore, the data obtained in the FPL calibration phase of the measurement may be used to independently estimate changes of the middle ear impedance. The same instrument was also used in a ground experiment to detect ICP changes associated with body-tilt in normal-hearing subjects. Experimental data from both experiments are shown to discuss strengths and weaknesses of the proposed combined method.